

CERTIFICATE OF COMPLIANCE

P/N 4370-010020

Mold # 33

Lot # 060507

Operator: 0467

Date: 05/07/06

Mod Block Vessels

Sample tubes, hinged caps, 70 mL, 400/pk

This vessel has been tested for volume at the 50mL, volume marking and is accurate to ± 0.25 mL.

All samples were certified using a 4 decimal place balance, which is calibrated with Class A NIST traceable weights.

For questions or comments please call 1-800-878-7654 in the USA or 31 20 638 05 07 in Europe.

CPI International